

CE 2007

USER'S GUIDE
for
MIDDLETON SOLAR
EQ08-S, EQ08-SE
SECONDARY STANDARD PYRANOMETER

Date: Sept. 2015

Version: 1.9



Middleton Solar, Australia.

© copyright 2015

www.middletonsolar.com

The contents of this manual are subject to change without prior notice.

CONTENTS		page
1	General	1
2	Construction	1
3	Installation	1
4	Maintenance	2
5	Calibration	3
6	Technical Specification	3
7	Spare Parts	4
Appendix A: Directional Response		5
Appendix B: Sensor Temperature vs Thermistor Resistance		6

1 GENERAL

The EQ08-S Pyranometer is for measuring total global solar irradiance on a plane surface. It exceeds the requirements for both an ISO9060 Secondary Standard Pyranometer and a WMO High Quality Pyranometer. The EQ08-SE version has an in-built signal amplifier.

2 CONSTRUCTION


The EQ08-S series is machined from anodised marine-grade aluminium to provide a durable corrosion-resistant instrument. It has a blackened thermal sensor shielded by two glass domes. The sensor incorporates a precision thermoelectric element that generates a passive microvolt output proportional to the intensity of the incident radiation. It is equally receptive to all radiation wavelengths and therefore its response to solar radiation is non-selective spectrally. The sensor has been designed to give exceptionally low directional error without compromise to signal strength or response time.

3 INSTALLATION

3.1 SITE SELECTION

For global radiation measurement, select a site where obstructions do not exceed 5° of elevation, in the path followed by the sun, between earliest sunrise and latest sunset during the year. Avoid positioning the instrument near light coloured objects (eg. painted walls) which can cause errors due to reflection of radiation onto the instrument; also avoid siting near sources of artificial light.

3.2 MOUNTING



The EQ08-S should be mounted on a rigid level surface. The instrument has a central M10 x 1.5p hole in its base. The mounting knob supplied should be screwed into this hole and the instrument mounted, as shown above, using the washer & nut provided. Adjust the height of the feet so the circular level is centred. Secure the knob & nut finger-tight only (do not over-tighten).

Alternatively the EQ08-S can be fastened using the two additional mount holes in the base. These holes are M4 x 0.7p, 65mm centre to centre.

3.3 ELECTRICAL CONNECTION, EQ08-S

The EQ08-S has a shielded output cable. The cable shield can be grounded at the measurement end.

The cable cores are:

red = output +ve (typically $13\mu\text{V}$ per W/m^2)

blue = output -ve

(option: yellow & green) = sensor temperature thermistor (see Appendix A))

The measurement equipment should range up to 50mV and have an input impedance of at least $1\text{M}\Omega$.

3.4 ELECTRICAL CONNECTION, EQ08-SE

The EQ08-SE is supplied with a 4-core lead:

red = supply +V (5 to 15VDC, < 6mA), blue = supply 0V

yellow = output +ve (1mV per W/m^2), green = output -ve

(OPTION: white & black = sensor temperature thermistor, see Appendix A)

The EQ08-SE can be powered from a 6V, 9V, or 12V battery. Alternatively it can be connected to a small power supply or datalogger. The typical fullscale output is <2V. The measuring equipment impedance should exceed $3\text{K}\Omega$. The settling time of the amplifier is 50ms from power-up.

Avoid ground loop induced interference in your measurement setup by ensuring there is only one ground point for the sensor and measurement system.

4 MAINTENANCE

Keep the outer glass dome clean. Use only water and mild detergent to gently wash the surface. If the outer dome is cracked or pitted it must be replaced. Periodically inspect the sensor plate to ensure that the black surface is not faded or flaking.

The EQ08-S sensor is hermetically sealed and contains a desiccant sachet to ensure that the interior remains dry and free from condensation. The desiccant is orange silica gel (non-toxic). The desiccant should be replaced when the instrument undergoes periodic calibration. The silica gel changes from orange to clear if moisture has entered the instrument.

To gain access to the dessicant sachet remove the three socket screws that hold the body to the top (it is not necessary to remove the shade disk). When refitting the body take care to locate the O-ring in its groove before securing the three socket screws.

5 CALIBRATION

Each pyranometer is individually calibrated during manufacture.

It is recommended that the instrument calibration be checked annually.

Two International Standards for pyranometer calibration are:

- ISO 9846: 1993(E), Calibration of a pyranometer using a pyrliometer.
- ISO 9847: 1992(E), Calibration of field pyranometers by comparison to a reference pyranometer.


6 TECHNICAL SPECIFICATION

Performance Specification	ISO9060 Secondary Standard (WMO High Quality)	EQ08-S/E (typical)
Response time (to 95%)	< 15s	14s
Zero off-set: a) 200 W.m ⁻²	7 W.m ⁻² (ventilated)	< 3 W.m ⁻² (ventilated) < 5 W.m ⁻² (unventilated)
b) 5K.h ⁻¹	± 2 W.m ⁻²	< ± 1 W.m ⁻²
Non-stability (per year)	± 0.8%	< - 0.5%
Non-linearity (100-1000W.m ⁻²)	± 0.5%	< ± 0.5%
Directional response (w.r.t. 1000 W.m ⁻² , 30-80°)	± 10 W.m ⁻²	± 5 W.m ⁻²
Spectral selectivity (0.3 to 3µm)	± 2%	< ±2%
Temperature response (for 50K interval)	± 2%	± 0.5%
Tilt response (0-90°)	± 0.5%	< ± 0.2%

General Specification	
viewing angle	2π steradians
irradiance	0 - 4000W/m ²
spectral range	300 - 3000nm (nominal) 305 - 2850nm (50% points)
sensitivity (typical)	EQ08-S: 13 µV/W.m ⁻² EQ08-SE: 1.0 mV/W.m ⁻²
signal resolution	< 0.5 W/m ²
impedance	EQ08-S: 40Ω; EQ08-SE: 65Ω
power requirement (EQ08-SE)	5-15VDC; < 6mA
operating temperature	-35 to +60°C
bubble level resolution	0.1°
level adjustment	one fixed foot, two adjustable feet
desiccant	orange silica gel (non-toxic, self-indicating)
mounting method	central M10 hole in base, plus pair M4 holes on 65mm P.C.D.
output lead	6m, with connector at instrument end


7 SPARE PARTS

Spare parts may be ordered from the manufacturer or through an approved distributor. For your convenience the part name and number is shown below. Please quote both when ordering. It is also important when ordering parts to include the Serial Number of the instrument, this is inscribed on the identification label of the unit.



note: part numbers valid for S/No.4905 onwards

Appendix A: DIRECTIONAL RESPONSE



Appendix B: SENSOR TEMPERATURE vs THERMISTOR RESISTANCE
For instruments fitted with optional sensor temperature output

YSI 44031 Thermistor (accuracy = $\pm 0.2^{\circ}\text{C}$)

Temperature ($^{\circ}\text{C}$)	Resistance (Ω)	Temperature ($^{\circ}\text{C}$)	Resistance (Ω)
-30	135,200	15	15,130
-29	127,900	16	14,500
-28	121,100	17	13,900
-27	114,600	18	13,330
-26	108,600	19	12,790
-25	102,900	20	12,260
-24	97,490	21	11,770
-23	92,430	22	11,290
-22	87,660	23	10,840
-21	83,160	24	10,410
-20	78,910	25	10,000
-19	74,910	26	9605
-18	71,130	27	9227
-17	67,570	28	8867
-16	64,200	29	8523
-15	61,020	30	8194
-14	58,010	31	7880
-13	55,170	32	7579
-12	52,480	33	7291
-11	49,940	34	7016
-10	47,540	35	6752
-9	45,270	36	6500
-8	43,110	37	6258
-7	41,070	38	6026
-6	39,140	39	5805
-5	37,310	40	5592
-4	35,570	41	5389
-3	33,930	42	5193
-2	32,370	43	5006
-1	30,890	44	4827
0	29,490	45	4655
1	28,150	46	4489
2	26,890	47	4331
3	25,690	48	4179
4	24,550	49	4033
5	23,460	50	3893
6	22,430	51	3758
7	21,450	52	3629
8	20,520	53	3504
9	19,630	54	3385
10	18,790	55	3270
11	17,980	56	3160
12	17,220	57	3054
13	16,490	58	2952
14	15,790	59	2854